

I have a Medtronic Defibrillator which is under advisory. Rapid battery depletion may occur.

See test instructions on other side.

Doctor: If abnormal test parameters:

MONITOR IN RESUSC ROOM and call:

705.749.8419 (0800-1600 Mon-Fri)

705.740.7177 (after hrs)

Test Procedure:

Apply magnet over device for maximum 5 seconds.

If NO AUDIBLE TONE, device has FAILED.

Continue to monitor and contact clinic.

Peterborough Regional Device Clinic